

---

## DECLARATION OF CONFORMITY

---

**COMPUR**  
Monitors

**Compur Monitors GmbH & Co.KG**  
**Weißenseestraße 101**  
**D 81539 München**

as the manufacturer hereby declares, that the

Gasalarm- and Monitoring Device  
**Monitox plus Type 5306 500**

complies with the essential requirements of the following directives and has been tested according to European standards:

1. Directive **89/336/EC**

**EN 50081-1**

**EN 55022**

**EN 50082-2**

**EN 55024**

2. Directive **94/9/EC**

**EN 50014 : 1997+A1-A2**

**EN 50020 : 1994**

EC Type Examination Certificate: DMT 02 ATEX E 033

Notified Body: DMT / 0158

Munich, 03-20-2003



Dr. H. Schmidt

---

COMPUR Monitors GmbH & Co.KG  
POB 900147  
D-81501 München  
DIN EN ISO 9001:2000 certified

Phone: ++49 89 62038268  
Internet <http://www.compur.com>  
e-mail: [compurmonitors@t-online.de](mailto:compurmonitors@t-online.de)

General Management:  
Dipl.-Ing. B. Rist  
Dr. H. Schmidt